



T:SLIM X2 INSULIN PUMP TRAINING Supply Checklist





t:slim X2 Insulin Pump Supplies



t:slim X2 insulin pump box



t:slim X2 insulin pump accessories box



2-3 cartridges with syringes and needles
Note: All located in the cartridge box.



2-3 infusion sets



Compatible smartphone with the t:connect mobile app installed

Additional Supplies



Full vial of Humalog or
NovoLog U-100 insulin



Alcohol swabs or
skin-prep wipes

Glucose Monitoring Supplies* (all may not apply)



Dexcom G6 CGM

Note: An active sensor can pair to your insulin pump during training.



Dexcom G7 CGM

Note: An active sensor can pair to your insulin pump during training. See below for important CGM compatibility information.



FreeStyle Libre 2 Plus Sensor

Note: Bring an unopened sensor to training to pair with your insulin pump. See below for important CGM compatibility information.



Blood glucose meter
with test strips

CGM Compatibility



Your Dexcom G7 sensor is compatible if the side of the box has a white line below the LBL number.



Your Tandem t:slim X2 insulin pump is compatible with FreeStyle Libre 2 Plus Sensor only.



Dexcom G7

For questions about compatible G7 sensors, visit tandemdiabetes.com/G7support. To request replacement G7 sensors, call (844) 607-8398.



FreeStyle Libre 2 Plus Sensor

For questions about compatible sensors, visit tandemdiabetes.com/AbbottSupport. To request replacement FreeStyle Libre 2 Plus Sensors, call (888) 963-8038.

*Glucose monitoring supplies sold separately.

Important Safety Information: RX ONLY. The t:slim X2 insulin pump with interoperable technology is intended for the subcutaneous delivery of insulin, at set and variable rates, for the management of diabetes mellitus in people requiring insulin. The pump is able to reliably and securely communicate with compatible, digitally connected devices, including automated insulin dosing software, to receive, execute, and confirm commands from these devices. The pump is intended for single patient use. The pump is indicated for use with NovoLog or Humalog U-100 insulin. The pump is indicated for use in individuals 6 years of age and greater. Users of the pump must: be willing and able to use the insulin pump and all other system components in accordance with their respective instructions for use; test blood glucose levels as recommended by their healthcare provider; demonstrate adequate carb-counting skills; maintain sufficient diabetes self-care skills; see healthcare provider(s) regularly; and have adequate vision and/or hearing to recognize all functions of the pump, including alerts. The t:slim X2 pump must be removed before MRI, CT, or diathermy treatment. Visit tandemdiabetes.com/safetyinfo for additional important safety information.

© 2024 Tandem Diabetes Care, Inc. All rights reserved. Tandem Diabetes Care, the Tandem logo, and t:slim X2 are either registered trademarks or trademarks of Tandem Diabetes Care, Inc. in the United States and/or other countries. Dexcom, Dexcom G6, Dexcom G7, and any related logos and design marks are either registered trademarks or trademarks of Dexcom, Inc. in the United States and/or other countries. The sensor housing, FreeStyle, Libre, and related brand marks are marks of Abbott and used with permission. All third-party marks are the property of their respective owners. ML-1012727_D



TANDEM
Diabetes Care

(877) 801-6901
tandemdiabetes.com

FOR USE IN THE
UNITED STATES ONLY

